

# ***CDC Hormone Standardization Program (CDC HoSt)***

## **Certified Estradiol Assays**

From 2019 Q4 and forward

(UPDATED 07/2021)

**CDC HoSt Programs started quarterly certification from November 2019 shipment.**

See [https://www.cdc.gov/labstandards/pdf/hs/CDC\\_Certified\\_Estradiol\\_Procedures-508.pdf](https://www.cdc.gov/labstandards/pdf/hs/CDC_Certified_Estradiol_Procedures-508.pdf) for previous list of certified assays.

- The following assays have successfully met the performance criteria of  $\pm 12.5\%$  mean bias (for samples  $>20$  pg/mL) and  $\pm 2.5$  pg/mL absolute bias (for samples  $\leq 20$  pg/mL) when compared to the CDC reference measurement procedure for estradiol for 80% of samples.
- It is not the intent of the CDC HoSt Program to certify each lot of reagents. Participants are awarded certificates for successfully meeting bias criteria using specific methods that consist of different reagent lots and calibrator lots.
- Analytical performance in CDC HoSt Program is assessed using human serum. The measurement procedures may have different accuracy and precision with other specimen types, such as plasma.
- Certification is valid for one quarter from the certification date. It is the responsibility of the participant to ensure that the results of the assay remain consistent, between lots, and over the measurement range reported.
- The analytical performance evaluation used in certification is for testing performed in patient care. Therefore, this certification does not imply suitability of a participant as a calibration laboratory or the procedure as a metrological reference measurement procedure.

Each table includes information about certified assays including participant name, measurement principle, assay identifier, assay measurement range, certification measurement range, certification date, individual samples pass rate, and contact information.

“Assay identifier” is an internal code used by the participant to represent the assay used for certification.

“Assay Measurement range” is the assays’ reported analytical measurement range (AMR) and is not the certification range.

“Certification Measurement Range” is the concentration range the of the samples used for HoSt certification.

“Certification date” includes historical certification information and gaps between years do not always indicate the assay’s failure to meet certification criteria.

“Individual samples pass rate” is the percentage of individual samples out of the 40 provided that met the certification criteria. This information was provided on the website starting June 2019.

**Table 1: Currently Certified Assays including their certification history**

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (pg/mL)	Certification Measurement Range (pg/mL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>ARUP Laboratories</b> Salt Lake City, UT	LC/MS/MS <sup>+</sup>	Estradiol by TMS	2 - 1000	3.05 - 230 2.55 - 230	Q1 2021 Q4 2020	95 92	Canary Tennison <a href="mailto:canary.tennison@aruplab.com">canary.tennison@aruplab.com</a> 801-583-2787 x2893
<b>Brigham Research Assay Core (BRAC) Laboratory at Harvard Medical School</b> Boston, MA	LC/MS/MS <sup>+</sup>	Serum Estradiol	1.00 - 500.00 (& higher than 500 pg/mL with dilution)	2.70 - 216 2.55 - 230 2.55 - 230 2.55 - 230 2.55 - 230 3.40 - 268	Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019	98 98 90 95 92 92	Dr. Shalender Bhasin <a href="mailto:SBHASIN@PARTNERS.ORG">SBHASIN@PARTNERS.ORG</a> (617)525-9040  Liming Peng <a href="mailto:Lpeng2@partners.org">Lpeng2@partners.org</a> (617)525-9048
<b>Centre Hospitalier Universitaire de Liège</b> Sart-Tilman, Belgium	LC/MS/MS <sup>+</sup>	25(OH)D3+25(OH)D2+C3-epimer+24,25(OH)2D in serum and plasma	5 - 1156	6.21 - 216	Q2 2020	85	Etienne Cavalier <a href="mailto:etienne.cavalier@chu.ulg.ac.be">etienne.cavalier@chu.ulg.ac.be</a> +3243667692
<b>Clinical Chemistry Branch CDC</b> Atlanta, GA	LC/MS/MS <sup>+</sup>	Total Estradiol in Serum (1036)	1.72 - 17,100	3.01 - 230 1.75 - 230 1.75 - 230 1.75 - 230 1.75 - 230 2.55 - 216	Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019	100 100 100 100 98 98	Lumi Duke, MS <a href="mailto:LDuke@cdc.gov">LDuke@cdc.gov</a> (770)488-4126
<b>Covance Central Laboratory Services</b> Indianapolis, IN	LC/MS/MS <sup>+</sup>	Total Estradiol in Serum (E2)	0.50 - 4,000	2.50 - 230 2.50 - 230 2.70 - 230 2.55 - 268 2.55 - 268 2.55 - 268	Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019	100 100 100 100 100 90	Cristina Hedin, MS Covance Central Laboratory Services <a href="mailto:Cristina.Hedin@covance.com">Cristina.Hedin@covance.com</a> 317-273-7842
<b>LabCorp</b> Burlington, NC	LC/MS/MS <sup>+</sup>	Estradiol in Serum	2.5 - 5000	2.55 - 230 2.55 - 230 2.55 - 230 2.55 - 230	Q1 2021 Q4 2020 Q3 2020 Q2 2020	90 90 88 85	Hema Ketha <a href="mailto:Kethah@LabCorp.com">Kethah@LabCorp.com</a> 336-436-3102

Participant	Measurement Principle	Assay Identifier	Assay Measurement Range (pg/mL)	Certification Measurement Range (pg/mL)	Certification Date (active for 1 quarter)	Individual Samples Pass Rate (%)	Participant's Contact Information
<b>LabCorp</b> Calabasas Hills, CA	LC/MS/MS <sup>†</sup>	#500108 Estradiol, LC/MS (Endocrine Sciences)	1 - 500 (1 to 5,000 with validated dilution)	2.55 - 230 2.55 - 230 2.55 - 230 2.55 - 230 2.55 - 268 3.40 - 268	Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019	98 98 90 88 90 90	Dr. Brett Holmquist <a href="mailto:holmqub@labcorp.com">holmqub@labcorp.com</a> (818) 867-1362  Dr. Kelly Chun <a href="mailto:chunk@labcorp.com">chunk@labcorp.com</a> (818) 867-1358
<b>LabCorp</b> Spokane, WA	LC/MS/MS <sup>†</sup>	ESTRADIOL (LCMSMS)	2.5 - 625 (2.5 - 5000 with validated dilution)	3.95 - 230 3.95 - 230 4.97 - 230 6.21 - 230 6.21 - 230 5.90 - 216	Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019	85 90 88 95 98 88	Carissa Schmitz MLS(ASCP)CM <a href="mailto:Schmic4@LabCrop.com">Schmic4@LabCrop.com</a> (509) 755-8358
<b>Mayo Clinic</b> Rochester, MN	LC/MS/MS <sup>†</sup>	Estradiol	10 - 600	11.4 - 230 11.7 - 230 11.8 - 230 17.1 - 230 17.1 - 216 16.6 - 230	Q1 2021 Q4 2020 Q3 2020 Q2 2020 Q1 2020 Q4 2019	100 100 95 95 95 92	Sue Reicks <a href="mailto:reicks.sue@mayo.edu">reicks.sue@mayo.edu</a>
<b>Roche Diagnostics GmbH</b> Penzberg, Germany	LC/MS/MS <sup>†</sup>	Total Estradiol in Serum and Plasma	5 - 5000	9.64 - 230 9.64 - 230	Q1 2021 Q4 2020	92 92	Tobias Santner <a href="mailto:tobias.santner@roche.com">tobias.santner@roche.com</a> +49 8856 605259

<sup>†</sup> LC/MS/MS – Liquid Chromatography Tandem Mass Spectrometry